SMART-TRIAL	**Review Record**	Doc. #: 15005-ST-0222

Organizer:	Laszlo Czegledi	**Subject:**	Record for approving SMART-TRIAL 2.25 is ready for release
Scribe:	Laszlo Czegledi		
Date:	01-Jul-2022		

Acceptance signatures:

Discipline or Role	*Print Name*	*Signature*	*Date*
Organizer	Laszlo Czegledi		01-Jul-2022
Approver	Amalie Randers Rosenlund		01-Jul-2022
Product Owner	Pall Johannesson		01-Jul-2022

5 This Document is the property of SMART-TRIAL ApS. The data contained herein, in whole or in part, may not be duplicated, used or disclosed without written permission from SMART-TRIAL ApS.

Printed copies must be verified for correctness against the approved QMS version

TABLE OF CONTENTS

1 PARTICIPANTS 3
2 SUMMARY OF DISCUSSIONS AND KEY DECISIONS 3
3 REQUIRED FOLLOW-UP REVIEWS 3
4 DETAILED REVIEW INPUT AND ACTION ITEMS 3

Printed copies must be verified for correctness against the approved QMS version

1 Participants

Laszlo Czegledi – Software developer for SMART-TRIAL v2.25
Amalie Randers Rosenlund – QA responsible for SMART-TRIAL v2.25 and reviewer
Pall Johannesson – Product owner for SMART-TRIAL v2.25

2 Summary of discussions and key decisions

With all documents reviewed and signed as per 15005-ST-0221, SMART-TRIAL 2.25 is ready for release with known defects with high severity identified:

- SMARTTRIAL-9364 - Multiple choice question shows up in the form even though it has a local show rule that should not allow it to be shown

Continually monitored on JIRA

One task was completed after the completion of the formal verification:

- Export fails when form name contains colon

(Full list found at https://jira.medei.co/issues/?jql=project%20%3D%20SMARTTRIAL%20AND%20fixVersion%20%3D%202.25%20AND%20labels%20%3D%20post-verification)

The impact of these changes was deemed to not require formal verification and validation process. Instead, the changes were tested locally. These tests were successful.

3 Required follow-up reviews

Not required

4 Detailed review input and action items

None

Printed copies must be verified for correctness against the approved QMS version